

LIFETIME LIMITED WARRANTY

Your Drive brand product is warranted to be free of defects in materials and workmanship for the lifetime of the original consumer purchaser.

This device was built to exacting standards and carefully inspected prior to shipment. This Lifetime Limited Warranty is an expression of our confidence in the materials and workmanship of our products and our assurance to the consumer of years of dependable service.

In the event of a defect covered by this warranty, we will, at our option, repair or replace the device.

This warranty does not cover device failure due to owner misuse or negligence, or normal wear and tear. The warranty does not extend to non-durable components, such as rubber accessories, casters, and grips, which are subject to normal wear and need periodic replacement.

If you have a question about your Drive device or this warranty, please contact an authorized Drive dealer.



Port Washington, NY 11050

MADE IN CHINA

© 2002 Medical Depot, Inc. All rights reserved.

Drive is a trademark of Medical Depot, Inc.

Port Washington N.Y. 11050 USA Made in China



Item #10233 Shown

Clever-Lite LS

Instructions & Warranty

www.drivemedical.com



Assembly Instructions:

Do not attempt to adjust or operate walker without carefully reading all instructions in this manual.

- 1** First remove all contents from carton: backrest, handle with tube, Walker frame and basket.
- 2** Open Frame and slide padded backrest into frame brackets making sure snap buttons protrude through holes.
- 3** Unscrew tightening knob from handle and insert handle tube into frame at desired height lining up holes appropriately. Slide bolts through holes from the inside and screw on tightening knob until secure.

Safety Precautions:

- Please make sure the Walker is fully opened, seat is down and the folding bar under seat is locked in the down position.
- Be sure that both knobs are fully tightened and are at equal heights. Test Rollator for stability before applying full weight to hand grips.
- Test the push down locks before operating the Walker. When wheels are locked (by sitting or having weight applied by pushing down on handles), rear wheels should prevent wheels from turning. If the wheels turn while locks are locked, please take unit back to dealer for adjustment. When weight is released, the wheels should turn freely.
- Weight will lock push locks by sitting. Do not use seat when unit is on an incline or uneven ground.
- Replace tires when significantly worn or damaged.
- Do not have anyone push you while you are seated on the Walker. **This is a walking aid only and is not to be used as a transportation device.**

Push Down Lock Operation:

Always engage and disengage wheel locks on both sides simultaneously.

To Slow Down or Come to a Temporary Stop — Push down slowly by applying weight on the handles until unit slows down or comes to a temporary stop.

To Temporarily Lock Wheels in Position — Once seated, rear wheels will automatically lock until weight is removed.

To Release Wheel Locks — Remove weight until wheels start moving freely.